

CALIBRATION CERTIFICATE

for

Kustom LASERwitness Lite

Mobile Speed Enforcement System

Customer: Dorset Police

Serial Number: LW06273

Certificate Number: 7364

Date of Calibration: 21 September 2023

Next Calibration: 20 September 2024

Place of Calibration: 14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

Calibration Results:

Checksum - CC85
Oscillator Frequency - 19.999948 MHz
Self Test Sequence - Pass
External Controls - Pass
Aiming - Vertical - Pass
Horizontal - Pass
Eye Safety - Pass

Range ± 0.5 ft

Actual	156.3 ft	356.6 ft	488.4 ft
Indicated	156.1 ft	356.7 ft	488.7 ft

Speed ± 1 MPH


Actual	23 MPH	74 MPH	-31 MPH	-49 MPH	59 MPH
Indicated	23 MPH	74 MPH	-31 MPH	-49 MPH	59 MPH

The sample measurements in the calibration certify this device for use to the full extent of the instrument specification Range 10 ft - 4500 ft, Speed 0 mph - 200 mph

The Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that if any parts were replaced during the calibration process that they are identical to those specified for use in the approved device. This device conforms with the device that is approved.

For Truvelo UK



L Korcz
Technical Signatory

TRUVELO (UK) LTD