CALIBRATION CERTIFICATE

for

Kustom LASERwitness Lite

Mobile Speed Enforcement System

Customer:	Dorset Police	**Serial Number:**	LW06274
		Certificate Number:	5390
Date of Calibration:	14 September 2021	**Next Calibration:**	13 September 2022
Place of Calibration:	14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG		

Calibration Results:

Checksum		- CC85
Oscillator Frequency		- 19.999633 MHz
Self Test Sequence		- Pass
External Controls		- Pass
Aiming	Vertical	- Pass
	Horizontal	- Pass
Eye Safety		- Pass

Range ± 0.5 ft					
Actual	156.3 ft	284.5 ft	488.4 ft		
Indicated	156.0 ft	284.3 ft	488.6 ft		

Speed ± 1 MPH					
Actual	27 MPH	71 MPH	-33 MPH	-45 MPH	53 MPH
Indicated	27 MPH	71 MPH	-33 MPH	-45 MPH	53 MPH

This is to certify that the Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that any parts replaced during the calibration process are identical to those used in the type approved device.

For Truvelo UK

L Korcz
Technical Signatory

TRUVELO (UK) LTD
Tel:020 8847 4400 email:info@truvelouk.com