

# CALIBRATION CERTIFICATE

for

## Kustom LaserCam 4 Speed Meter

**Customer:** Dorset Police **Serial Number:** LE1011

**Certificate Number:** 6656

**Date of Calibration:** 01 December 2022 **Next Calibration:** 30 November 2023

**Place of Calibration:** 14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

### Calibration Results:

Oscillator Frequency - 19.999788 MHz  
Self Test Sequence - Pass  
External Controls - Pass  
Aiming Vertical - Pass  
Horizontal - Pass  
Eye Safety - Pass

Range ± 0.2 m			
Actual	47.64 m	108.69 m	148.86 m
Indicated	47.6 m	108.7 m	148.9 m

Speed ± 1 MPH					
Actual	51 MPH	73 MPH	-26 MPH	-33 MPH	58 MPH
Indicated	51 MPH	73 MPH	-26 MPH	-33 MPH	58 MPH

This is to certify that the Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that any parts replaced during the calibration process are identical to those used in the type approved device.

For Truvelo UK

L Korcz  
Technical Signatory

**TRUVELO (UK) LTD**