CALIBRATION CERTIFICATE

for

Kustom LaserCam 4 Speed Meter

Customer:	Dorset Police	**Serial Number:**	LE03365
		Certificate Number:	7901
Date of Calibration:	11 April 2024	**Next Calibration:**	10 April 2025
Place of Calibration:	14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG		

Calibration Results:

Checksum	-	CD748BAF
Oscillator Frequency	-	19.999809 MHz
Self Test Sequence	-	Pass
External Controls	-	Pass
Aiming - Vertical	-	Pass
Aiming - Horizontal	-	Pass
Eye Safety	-	Pass

Range ± 0.2 m			
Actual	47.64 m	108.69 m	148.86 m
Indicated	47.6 m	108.7 m	148.9 m

Speed ± 1 MPH					
Actual	44 MPH	30 MPH	-21 MPH	52 MPH	-73 MPH
Indicated	44 MPH	30 MPH	-21 MPH	62 MPH	-73 MPH

The sample measurements in the calibration certify this device for use to the full extent of the instrument specification Range 3 m - 750 m, Speed 0 mph - 200 mph

The Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that if any parts were replaced during the calibration process that they are identical to those specified for use in the approved device. This device conforms with the device that is approved.

For Truvelo UK

L Korez
Technical Signatory

TRUVELO (UK) LTD

Tel:020 8847 4400 email:info@truvelouk.com