

CALIBRATION CERTIFICATE

for

Kustom LASERwitness Lite Mobile Speed Enforcement System

Customer: Dorset Police

Serial No: LW06272

Certificate No: T3207

Date of Calibration: 2 December 2014

Next Calibration: 1 December 2015

Checksum: CC85

Pulse repetition freq.: 200 Hz

Self test sequence: PASS

External controls: PASS

Aiming:- Vertical: CORRECT

Horizontal: CORRECT

Eye safety: PASS

Range (+/- 1ft)

Actual: 156.3 284.5 488.4

Indicated: 156.3 284.5 488.3

Speed (+/- 1ft)

Actual: (1) 24 (2) -35 (3) 52 (4) -66 (5) 74

Indicated: (1) 24 (2) -35 (3) 52 (4) -65 (5) 74

Calibrated by:

Calibration Signature

This is to certify that the Laser speedmeter detailed above has been calibrated to Home Office requirements, using test instruments calibrated to standards traceable to NPL.

We certify that any parts replaced during the calibration process are identical to those used in the type approved device.

For Truvelo (UK) Ltd

Service Manager

TRUVELO (UK) LTD

Tel:020 8847 4400

email:info@truelouk.com