

CALIBRATION CERTIFICATE

for
Kustom LASERwitness Lite
Mobile Speed Enforcement System

Customer: Dorset Police **Serial Number:** LW06273

Certificate Number: T5076

Date of Calibration: 14 December 2017 **Next Calibration:** 13 December 2018

Place of Calibration: 14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

Calibration Results:

Checksum - CC85
Oscillator Frequency - 19.999784 MHz
Self Test Sequence - Pass
External Controls - Pass
Aiming - Vertical - Pass
 - Horizontal - Pass
Eye Safety - Pass

Range ± 0.5 ft

Actual	156.3 ft	284.5 ft	488.4 ft
Indicated	156.1 ft	284.3 ft	488.7 ft

Speed ± 1 MPH

Actual	74 MPH	-72 MPH	59 MPH	-31 MPH	26 MPH
Indicated	74 MPH	-72 MPH	59 MPH	-31 MPH	26 MPH

This is to certify that the Laser speedmeter detailed above has been calibrated to Home Office requirements; test instruments calibrated to standards traceable to NPL.

We certify that any parts replaced during the calibration process are identical to those used in the type approved device.

For Truvelo UK


L Korcz
Technical Signatory

TRU-CAL-CER-005

TRUVELO (UK) LTD

Tel: 020 8847 4400 email: info@truvelouk.com

