

# TELE-TRAFFIC (UK) LIMITED

## CALIBRATION DATA SHEET

LTI 20-20 TruCAM

Certificate No 18- 0648

MANUFACTURED BY: LASER TECHNOLOGY INC.

MANUFACTURERS SERIAL NUMBER: TC 004891

UNITED KINGDOM SERIAL NUMBER: 449636TC

Visual Inspection: Pass  
Initialisation Test: Pass  
Display Test: Pass  
Firmware Version Check: Pass

### Distance Checks

No:1	25 Metres	Pass	Speed: 0 mph
No:2	75 Metres	Pass	Speed: 0 mph
No:3	100 Metres	Pass	Speed: 0 mph

Pulse Repetition Frequency Test 200.321 ppm nominal  
Output Period Specification: 200.221 to 200.421 Hz  
Measured Output Period: 200.320 Hz

### Speed Simulation Tests

Specification: +/- 1MPH or 2 KPH

Measured Test No:1	@ 20MPH	= 19
Measured Test No 2	@ 30MPH	= 29
Measured Test No 3	@ 50MPH	= 49
Measured Test No:4	@ 70MPH	= 69
Measured Test No:5	@ 80MPH	= 80

Scope Alignment Test: Pass  
(Horizontal and Vertical alignment correct)

### NOTES

Any parts or components that may have been replaced during the calibration procedure are certified as identical to those used in the device as type approved.

I CERTIFY THAT THE FOREGOING INTERNAL TESTS AND TEST DATA ARE CORRECT  
AND THAT THE INSTRUMENT ALIGNMENT IS CORRECT

TESTED BY: James Sample

DATE: : 24 July 2018  
CAL DUE: : 24 July 2019

TIME: 11:54 AM

# TELE-TRAFFIC (UK) LIMITED CALIBRATION DATA SHEET

LTI 20-20 TruCAM

Certificate No 17- 0485

MANUFACTURED BY: LASER TECHNOLOGY INC.

MANUFACTURERS SERIAL NUMBER: TC 004891

UNITED KINGDOM SERIAL NUMBER: 449636TC

Visual Inspection: Pass  
Initialisation Test: Pass  
Display Test: Pass  
Serial Port Check: Pass

#### Distance Checks

No:1	25 Metres	Pass	Speed: 0 mph
No:2	75 Metres	Pass	Speed: 0 mph
No:3	100 Metres	Pass	Speed: 0 mph

Pulse Repetition Frequency Test: 200.321 ppm nominal  
Output Period Specification: 200.221 to 200.421 Hz  
Measured Output Period: 200.321 Hz

#### Speed Simulation Tests

Specification: +/- 1MPH or 2 KPH

Measured Test No:1	@ 20MPH	= 19
Measured Test No 2	@ 30MPH	= 30
Measured Test No 3	@ 50MPH	= 50
Measured Test No:4	@ 70MPH	= 70
Measured Test No:5	@ 80MPH	= 80

#### Scope Alignment Test:

(Horizontal and Vertical alignment correct)

Pass

#### NOTES

Any parts or components that may have been replaced during the calibration procedure are certified as identical to those used in the device as type approved.

I CERTIFY THAT THE FOREGOING INTERNAL TESTS AND TEST DATA ARE CORRECT  
AND THAT THE INSTRUMENT ALIGNMENT IS CORRECT

TESTED BY: James Sample

DATE: : 12 May 2017  
CAL DUE: : 12 May 2018

TIME: 11:05 AM