## **CALIBRATION CERTIFICATE**

for

## **Kustom LASERwitness Lite**

**Mobile Speed Enforcement System** 

Customer:

**Dorset Police** 

Serial Number:

LW06274

**Certificate Number:** 

6406

**Date of Calibration:** 

20 September 2022

**Next Calibration:** 

19 September 2023

Place of Calibration:

14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

Calibration Results:

Checksum

- CC85

Oscillator Frequency

- 19.999624 MHz

Self Test Sequence

- Pass

External Controls

- Pass

Aiming

- Pass

Horizontal

Vertical

- Pass

Eye Safety

- Pass

Range ± 0.5 ft

Actual

156.3 ft

284.5 ft

488.4 ft

Indicated

156.2 ft

284.5 ft

488.6 ft

Speed ± 1 MPH

Actual

74 MPH

26 MPH

-38 MPH

52 MPH

-61 MPH

Indicated

74 MPH

26 MPH

-38 MPH

52 MPH

-61 MPH

This is to certify that the Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that any parts replaced during the calibration process are identical to those used in the type approved device.

For Truvelo UK

**L**Korcz

**Technical Signatory** 

TRUVELO (UK) LTD

TRU-CAL-CER-005 Tel:020 8847 4400

email:info@truvelouk.com